

(1) 22 ga. from each indicating point

This diagram represents a typical basic system. For more specific details and variations, please refer to the AlphaLINQTM 100 Series Installation & Operation Manual AWD183.



TYPICAL BLOCK WIRING DIAGRAM FOR SINGLE MASTER AlphaLINQTM 100 SERIES NURSE-CALL SYSTEM - without 3rd priority level for Code Calls or Smoke Detectors etc. This is a standard 2-priority level system.